PROJECT TITLE : ANALYTICAL INVESTIGATIONS

PERIOD COVERED : MARCH - APRIL 15, 1981

WRITTEN BY : Piadé-J-J. (JJP)

HPLC

The RI detector was installed on the HP 1084 and performance optimisation of the LC system carried out, requiring notably detector cell thermostatisation.

Detection limit for fructose was brought down to approximately 1 µg.

Radiocompounds

A recovery check by ^{14}C standards was carried out on the work-up for organic acids analysis by GC(1).

REFERENCE

(1) Memo March 25, 1981, from Gruring-D.

JJP/jig/APRIL 21 1981

00014426

Source: https://www.industrydocuments.ucsf.edu/docs/pghk0000